

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	General Medical Devices Thermotherapy device	
Name and address of the applicant	Hilotherm GmbH Wittumweg 38 88260 Argenbühl-Eisenharz GERMANY	
Name and address of the manufacturer	Hilotherm GmbH Wittumweg 38, 88260 Argenbühl-Eisenharz, GERMANY	
Name and address of the factory	Hilotherm GmbH Wittumweg 38, 88260 Argenbühl-Eisenharz, GERMANY	
Ratings and principal characteristics	Rated voltage	100 - 240 V _{AC}
	Rated frequency	47 - 63 Hz
	Rated input power	2.2 - 1.1 A
	Class of protection	II
	Mode of operation	continuously
	Type of applied part	BF
Trade mark (if any)	HILOTHERM	
Model/type Ref.	Hilotherm HTP1 Pro	
Additional information (if necessary)	Certificate DE 3 - 40550 issued 2016-06-21 is replaced by this version due to technical changes.	
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013 IEC 60601-1-11:2015	
as shown in the Test Report Ref. No. which forms part of this certificate	713174361	

This CB Test Certificate is issued by the National Certification Body

CB 096215 0007 Rev. 00

Date, 2020-07-27



(Paul Bielawski)

Page 1 of 1

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service