



Ref. Certif. No.

DE 3 - 41719

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	General Medical Devices Hypo-/hyperthermal device	
Name and address of the applicant	<b>Hilotherm GmbH</b> Wittumweg 38 88260 Argenbühl-Eisenharz GERMANY	
Name and address of the manufacturer	Hilotherm GmbH Wittumweg 38, 88260 Argenbühl-Eisenharz, GERMANY	
Name and address of the factory	Hilotherm GmbH Wittumweg 38, 88260 Argenbühl-Eisenharz, GERMANY	
Ratings and principal characteristics	Rated voltage	230 VAC, 50 Hz or 115 VAC, 60 Hz
	Power consumption	320 VA
	Protection class	I
	Mode of operation	continuously
	Degree of protection	IP20
	Type of applied part	B
Trademark / Brand (if any)	HILOTHERM	
Model/type Ref.	Hilotherm HT02	
Additional information (if necessary)	Certificate DE 3 - 41535 issued 2020-11-16 was replaced by this version due to technical changes.	
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1:2005/AMD2:2020 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013 IEC 60601-1-6:2010/AMD2:2020	
as shown in the Test Report Ref. No. which forms part of this certificate	713299657	

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 096215 0009 Rev. 00

Date, 2023-10-02

( Paul Bielawski )

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service